

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99922	0.780	± 1	± 6.6E-01 mm					
0 a 2	1.99923	0.770	± 1.4	± 7.0E-01 mm					
0 a 3	2.99926	0.740	± 1.8	± 7.0E-01 mm					
0 a 4	3.99928	0.720	± 2.2	± 7.0E-01 mm					
0 a 5	4.99928	0.720	± 2.6	± 7.0E-01 mm					
0 a 6	5.99929	0.710	± 3	± 6.1E-01 mm					
0 a 7	6.99931	0.690	± 3.4	± 6.1E-01 mm					
0 a 8	7.99933	0.670	± 3.8	± 6.1E-01 mm					
0 a 9	8.99932	0.680	± 4.2	± 6.1E-01 mm					
0 a 10	9.99935	0.650	± 4.6	± 6.1E-01 mm					
0 a 11	10.99967	0.640	± 5	± 6.1E-01 mm					
0 a 12	11.99965	0.610	± 5.4	± 6.1E-01 mm					
0 a 13	12.99961	0.620	± 5.8	± 6.1E-01 mm					
0 a 14	13.99960	0.590	± 6.2	± 6.1E-01 mm					
0 a 15	14.99956	0.580	± 6.6	± 6.1E-01 mm					