

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 1	0.99923	0.770	± 1	±	6.6E-01 mm						
0 a 2	1.99921	0.790	± 1.4	±	7.0E-01 mm						
0 a 3	2.99924	0.760	± 1.8	±	7.0E-01 mm						
0 a 4	3.99927	0.730	± 2.2	±	7.0E-01 mm						
0 a 5	4.99930	0.700	± 2.6	±	7.0E-01 mm						
0 a 6	5.99932	0.680	± 3	±	6.1E-01 mm						
0 a 7	6.99934	0.660	± 3.4	±	6.1E-01 mm						
0 a 8	7.99937	0.630	± 3.8	±	6.1E-01 mm						
0 a 9	8.99940	0.600	± 4.2	±	6.1E-01 mm						
0 a 10	9.99942	0.580	± 4.6	±	6.1E-01 mm						
0 a 11	10.99970	0.610	± 5	±	6.1E-01 mm						
0 a 12	11.99965	0.620	± 5.4	±	6.1E-01 mm						
0 a 13	12.99967	0.550	± 5.8	±	6.1E-01 mm						
0 a 14	13.99964	0.550	± 6.2	±	6.1E-01 mm						
0 a 15	14.99965	0.490	± 6.6	±	6.1E-01 mm						