

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	28.789 °C	29.3 °C	0.511 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.781 %HR	24.9 %HR	-3.881 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.006 %HR	45.6 %HR	-3.406 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	72.523 %HR	69.0 %HR	-3.523 %HR	± 5.0 %HR	± 1.9E+00 %HR