

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.042 °C	30.4 °C	0.358 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.478 %HR	37.5 %HR	2.022 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	51.2 %HR	0.845 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	71.1 %HR	0.448 %HR	± 3.0 %HR	± 1.9E+00 %HR