

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.687 °C	20.4 °C	-0.287 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.641 °C	29.5 °C	-0.141 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.070 °C	39.9 °C	-0.170 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.885 %HR	52.1 %HR	2.215 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.374 %HR	61.0 %HR	1.626 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.826 %HR	70.9 %HR	0.074 %HR	± 5.0 %HR	± 1.9E+00 %HR