

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
20 °C	20.687 °C	20.3 °C	-0.387 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.641 °C	29.3 °C	-0.341 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.070 °C	39.7 °C	-0.370 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	49.885 %HR	52.3 %HR	2.415 %HR	± 5.0 %HR	± 1.9E+00 %HR
60 %HR	59.374 %HR	63.1 %HR	3.726 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.826 %HR	74.1 %HR	3.274 %HR	± 5.0 %HR	± 1.9E+00 %HR