

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.957 °C	10.6 °C	0.643 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.536 °C	21.1 °C	0.564 °C	± 5.0 °C	± 4.0E-01 °C
30 °C	30.012 °C	30.3 °C	0.288 °C	± 5.0 °C	± 4.0E-01 °C
# 10 %HR	10.836 %HR	20.0 %HR	9.164 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	40.0 %HR	-10.284 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	80.366 %HR	70.0 %HR	-10.366 %HR	± 3.0 %HR	± 1.9E+00 %HR