

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.042 °C	29.1 °C	-0.942 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	31.0 %HR	0.634 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	51.0 %HR	0.645 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	73.0 %HR	2.348 %HR	± 5.0 %HR	± 1.9E+00 %HR