

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
30 °C	30.042 °C	29.6 °C	-0.442 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	27.0 %HR	-3.366 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	54.0 %HR	3.645 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	75.0 %HR	4.348 %HR	± 5.0 %HR	± 1.9E+00 %HR