

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.151 °C	15.4 °C	0.249 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.4 °C	-0.050 °C	± 1.0 °C	± 4.0E-01 °C
50 °C	50.368 °C	49.9 °C	-0.468 °C	± 1.0 °C	± 1.1E+00 °C
25 %HR	25.355 %HR	25.0 %HR	-0.355 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	48.0 %HR	-2.355 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	72.0 %HR	-3.253 %HR	± 5.0 %HR	± 1.9E+00 %HR