

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 52.65 μm          | 52.78 μm  | 0.13 μm       | ± 2.5 μm          | ±                           | 5.8E-02 μm |
| 155.23 μm         | 154.4 μm  | -0.83 μm      | ± 2.5 μm          | ±                           | 6.2E-02 μm |
| 257.12 μm         | 257.0 μm  | -0.12 μm      | ± 2.5 μm          | ±                           | 7.0E-02 μm |
| 511.72 μm         | 512.4 μm  | 0.68 μm       | ± 2.5 μm          | ±                           | 9.7E-02 μm |