

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.032 °C	30.1 °C	0.068 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	33.0 %HR	2.634 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	50.0 %HR	-0.355 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	68.0 %HR	-2.652 %HR	± 5.0 %HR	± 1.9E+00 %HR