

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.4 °C	0.393 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.035 °C	20.5 °C	0.465 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.5 °C	0.438 °C	± 1.0 °C	± 4.0E-01 °C