

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	14.670 °C	14.9 °C	0.230 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.133 °C	30.5 °C	0.367 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.558 %HR	32.3 %HR	2.742 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.558 %HR	66.8 %HR	-2.758 %HR	± 3.0 %HR	± 1.9E+00 %HR