

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.012 °C	29.5 °C	-0.512 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	32.0 %HR	1.737 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	51.0 %HR	0.716 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	68.0 %HR	-2.141 %HR	± 5.0 %HR	± 1.9E+00 %HR