

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.012 °C	30.0 °C	-0.012 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	33.0 %HR	2.737 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	52.0 %HR	1.716 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	70.0 %HR	-0.141 %HR	± 5.0 %HR	± 1.9E+00 %HR