

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.250 °C	29.0 °C	-0.250 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.556 %HR	28 %HR	-0.556 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.678 %HR	47 %HR	-1.678 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.407 %HR	75 %HR	-0.407 %HR	± 5.0 %HR	± 1.9E+00 %HR