

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.153 °C	29.6 °C	-0.553 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.936 %HR	32.7 %HR	2.764 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.854 %HR	49.6 %HR	-0.254 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.877 %HR	68.3 %HR	-2.577 %HR	± 3.0 %HR	± 1.9E+00 %HR