

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.153 °C	29.9 °C	-0.253 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.936 %HR	32.4 %HR	2.464 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.854 %HR	50.2 %HR	0.346 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.877 %HR	68.0 %HR	-2.877 %HR	± 3.0 %HR	± 1.9E+00 %HR