

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.153 °C	29.4 °C	-0.753 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.936 %HR	31 %HR	1.064 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.854 %HR	49 %HR	-0.854 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.877 %HR	71 %HR	0.123 %HR	± 5.0 %HR	± 1.9E+00 %HR