

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.153 °C	29.4 °C	-0.753 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.936 %HR	30 %HR	0.064 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.854 %HR	46 %HR	-3.854 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.877 %HR	67 %HR	-3.877 %HR	± 5.0 %HR	± 1.9E+00 %HR