

Certificado No. (Certificate #):61208Equipo (Instrument):Micrómetro de Exterior

Fecha de Calibración (Calibration Date):2025-09-23Marca (Brand):ACCUSIZE

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.0000 in	1.0000 in	0.0000 in	± 0.0001 in	±	3.4E-05 in
1.1000 in	1.1001 in	0.0001 in	± 0.0002 in	±	3.8E-05 in
1.2000 in	1.2001 in	0.0001 in	± 0.0002 in	±	5.2E-05 in
1.3000 in	1.3001 in	0.0001 in	± 0.0002 in	±	3.4E-05 in
1.4000 in	1.4000 in	0.0000 in	± 0.0002 in	±	5.4E-05 in
1.5000 in	1.5001 in	0.0001 in	± 0.0002 in	±	4.7E-05 in
1.6000 in	1.6001 in	0.0001 in	± 0.0002 in	±	5.2E-05 in
1.7000 in	1.7001 in	0.0001 in	± 0.0002 in	±	6.5E-05 in
1.8000 in	1.8001 in	0.0001 in	± 0.0002 in	±	5.8E-05 in
1.9000 in	1.9001 in	0.0001 in	± 0.0002 in	±	4.8E-05 in
2.0000 in	2.0000 in	0.0000 in	± 0.0002 in	±	4.2E-05 in