

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	25.049 °C	25.8 °C	0.751 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.611 %HR	49.0 %HR	-1.611 %HR	± 5.0 %HR	± 1.9E+00 %HR