

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.521 °C        | -25.1 °C  | 0.421 °C      | ± 1.1 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.171 °C         | 0.1 °C    | 0.271 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.635 °C          | 5.8 °C    | 0.165 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 9.907 °C          | 10.2 °C   | 0.293 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.149 °C         | 25.4 °C   | 0.251 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 29.852 °C         | 30.0 °C   | 0.148 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.010 °C         | 40.2 °C   | 0.190 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

