

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.521 °C	-25.0 °C	0.521 °C	± 1.1 °C	± 4.0E-01 °C
0 °C	-0.171 °C	0.2 °C	0.371 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.635 °C	6.0 °C	0.365 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.907 °C	10.2 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.3 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.852 °C	29.9 °C	0.048 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.010 °C	39.7 °C	-0.310 °C	± 0.5 °C	± 4.0E-01 °C

