

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.270 °C	-25.0 °C	0.270 °C	± 1.1 °C	± 4.0E-01 °C
0 °C	-0.291 °C	-0.1 °C	0.191 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.214 °C	5.4 °C	0.186 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.877 °C	10.3 °C	0.423 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.888 °C	25.1 °C	0.212 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.711 °C	30.0 °C	0.289 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.759 °C	40.2 °C	0.441 °C	± 0.5 °C	± 4.0E-01 °C

