

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.290 °C	-25.3 °C	-0.010 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	0.110 °C	0.2 °C	0.090 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.1 °C	0.107 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.998 °C	25.2 °C	0.202 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.112 °C	30.4 °C	0.288 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.240 °C	40.5 °C	0.260 °C	± 0.5 °C	± 4.0E-01 °C

