

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.290 °C        | -25.0 °C  | 0.290 °C      | ± 0.70 °C            | ± 4.0E-01 °C                |
| 0 °C              | 0.110 °C          | 0.0 °C    | -0.110 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 5.2 °C    | 0.207 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.007 °C         | 10.1 °C   | 0.093 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.998 °C         | 24.8 °C   | -0.198 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.112 °C         | 29.9 °C   | -0.212 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.240 °C         | 40.0 °C   | -0.240 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

