

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.521 °C	-25.7 °C	-0.179 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.171 °C	-0.3 °C	-0.129 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.635 °C	5.3 °C	-0.335 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.907 °C	9.7 °C	-0.207 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.149 °C	24.8 °C	-0.349 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.852 °C	29.7 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.010 °C	39.8 °C	-0.210 °C	± 0.5 °C	± 4.0E-01 °C

