

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.521 °C	-25.4 °C	0.121 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.171 °C	0.0 °C	0.171 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.635 °C	5.8 °C	0.165 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.907 °C	10.2 °C	0.293 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.149 °C	25.3 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.852 °C	30.1 °C	0.248 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.010 °C	40.2 °C	0.190 °C	± 0.5 °C	± 4.0E-01 °C

