

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.072 °C	30.0 °C	-0.072 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.069 %HR	25 %HR	-5.069 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	45 %HR	-5.028 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.926 %HR	71 %HR	-3.926 %HR	± 5.0 %HR	± 1.9E+00 %HR