

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.072 °C	29.0 °C	-1.072 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.069 %HR	24 %HR	-6.069 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	38 %HR	-12.028 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.926 %HR	59 %HR	-15.926 %HR	± 5.0 %HR	± 1.9E+00 %HR