

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
10 RPM	10.000 RPM	0.000 RPM	± 0.0060 RPM	±	4.0E-03 RPM
100 RPM	100.00 RPM	0.00 RPM	± 0.060 RPM	±	7.0E-03 RPM
6000 RPM	6000 RPM	0 RPM	± 4.0 RPM	±	5.8E-01 RPM