

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.5 RPM	0.5 RPM	0 RPM	± 0.001 RPM	±	4.0E-03 RPM
100 RPM	100 RPM	0 RPM	± 0.05 RPM	±	4.0E-03 RPM
1000 RPM	1000 RPM	0 RPM	± 0.51 RPM	±	7.0E-03 RPM
6000 RPM	6000 RPM	0 RPM	± 4.00 RPM	±	5.8E-01 RPM