

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.222 °C	14.9 °C	-0.322 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.3 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.4 °C	0.338 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.448 %HR	71.2 %HR	0.752 %HR	± 3.0 %HR	± 1.9E+00 %HR