

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.222 °C                  | 15.1 °C            | -0.122 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.349 °C                  | 25.3 °C            | -0.049 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 °C                      | 30.062 °C                  | 30.3 °C            | 0.238 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 70 %HR                     | 70.448 %HR                 | 71.3 %HR           | 0.852 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |

