

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.00 mmHg	50 mmHg	0.00 mmHg	± 3 mmHg	±	1.2E+00 mmHg
100.00 mmHg	100 mmHg	0.00 mmHg	± 3 mmHg	±	1.2E+00 mmHg
150.00 mmHg	150 mmHg	0.00 mmHg	± 3 mmHg	±	1.2E+00 mmHg
200.00 mmHg	200 mmHg	0.00 mmHg	± 3 mmHg	±	1.2E+00 mmHg
250.00 mmHg	250 mmHg	0.00 mmHg	± 3 mmHg	±	1.2E+00 mmHg
300.00 mmHg	300 mmHg	0.00 mmHg	± 3 mmHg	±	1.2E+00 mmHg