

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.925 °C	1.6 °C	-0.325 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.831 °C	7.5 °C	-0.331 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.877 °C	9.6 °C	-0.277 °C	± 0.5 °C	± 4.0E-01 °C