

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.925 °C	2.3 °C	0.375 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.831 °C	8.3 °C	0.469 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.877 °C	10.3 °C	0.423 °C	± 0.5 °C	± 4.0E-01 °C