

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.925 °C	2.3 °C	0.375 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.831 °C	8.0 °C	0.169 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.877 °C	10.1 °C	0.223 °C	± 0.5 °C	± 4.0E-01 °C