

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.262 °C	15.7 °C	0.438 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.6 °C	0.301 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.203 °C	30.5 °C	0.297 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.486 %HR	58.7 %HR	-0.786 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.458 %HR	67.6 %HR	2.142 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.885 %HR	74.2 %HR	4.315 %HR	± 5.0 %HR	± 1.9E+00 %HR