

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.262 °C	15.6 °C	0.338 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.6 °C	0.301 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.203 °C	30.7 °C	0.497 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.486 %HR	57.4 %HR	-2.086 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	65.458 %HR	65.5 %HR	0.042 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.885 %HR	73.1 %HR	3.215 %HR	± 3.0 %HR	± 1.9E+00 %HR