

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.173 °C	-29.7 °C	0.473 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.140 °C	-24.7 °C	0.440 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.242 °C	-14.9 °C	0.342 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.559 °C	-10.3 °C	0.259 °C	± 3.0 °C	± 4.0E-01 °C
2 °C	1.925 °C	2.0 °C	0.075 °C	± 3.0 °C	± 4.0E-01 °C
8 °C	7.831 °C	8.0 °C	0.169 °C	± 3.0 °C	± 4.0E-01 °C