

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
16 °C	15.903 °C	15.2 °C	-0.703 °C	± 1.0 °C	± 4.0E-01 °C
21 °C	20.125 °C	19.4 °C	-0.725 °C	± 1.0 °C	± 4.0E-01 °C
60 %HR	60.079 %HR	62 %HR	1.921 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.967 %HR	77 %HR	2.033 %HR	± 5.0 %HR	± 1.9E+00 %HR