

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	0.10 mm	0.10 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	0.50 mm	0.50 mm	0 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	1.00 mm	1.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	2.00 mm	2.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	5.00 mm	5.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	10.00 mm	10.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	15.00 mm	15.02 mm	20 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	20.00 mm	20.03 mm	30 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	25.00 mm	25.03 mm	30 µm	± 30 µm	± 6.1E+00 µm