

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
50.0 mmHg	50 mmHg	0.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
100.0 mmHg	100 mmHg	0.0 mmHg	± 5 mmHg	±	6.5E-01 mmHg
150.0 mmHg	150 mmHg	0.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
200.0 mmHg	200 mmHg	0.0 mmHg	± 5 mmHg	±	6.8E-01 mmHg
250.0 mmHg	250 mmHg	0.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg
299.0 mmHg	300 mmHg	1.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg