

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.0 mmHg	50 mmHg	1.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
100.0 mmHg	100 mmHg	0.0 mmHg	± 5 mmHg	±	6.5E-01 mmHg
151.0 mmHg	150 mmHg	-1.0 mmHg	± 5 mmHg	±	6.6E-01 mmHg
200.0 mmHg	200 mmHg	0.0 mmHg	± 5 mmHg	±	6.8E-01 mmHg
251.0 mmHg	250 mmHg	-1.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg
279.0 mmHg	280 mmHg	1.0 mmHg	± 5 mmHg	±	7.5E-01 mmHg