

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.072 °C         | 29.8 °C   | -0.272 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.069 %HR        | 32.9 %HR  | 2.831 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.028 %HR        | 52.0 %HR  | 1.972 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 75 %HR            | 74.926 %HR        | 72.3 %HR  | -2.626 %HR    | ± 3.0 %HR            | ± 1.9E+00 %HR               |