

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.072 °C	30.1 °C	0.028 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.069 %HR	32.5 %HR	2.431 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	51.8 %HR	1.772 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.926 %HR	72.5 %HR	-2.426 %HR	± 3.0 %HR	± 1.9E+00 %HR