

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.072 °C	30.3 °C	0.228 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.069 %HR	32.2 %HR	2.131 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	51.8 %HR	1.772 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.926 %HR	73.8 %HR	-1.126 %HR	± 3.0 %HR	± 1.9E+00 %HR